

ACS

**AUTOMATED CALIBRATION
AND REPAIR SERVICES**

AUTOMATED CALIBRATION SERVICES
UNIT 12, ACCESS 18
KINGS WESTON LANE
BRISTOL, BS11 8HT

Tel: 0117 938 6401 Fax: 0117 923 5374

Certificate No.

0262538

Certificate of Calibration

Date of Issue: **14-06-2023**

Page 1 of 3

Job No: ***9004/CEF VOUCHER 23**

Client: **RIS SVS
c/o CITY ELECTRICAL FACTORS LTD
20 HARVEST DRIVE SOUTH
LOWESTOFT IND. EST.
LOWESTOFT
SUFFOLK
NR33 7NJ**

Description: **Digital Multifunction Tester**

Manufacturer: **Kewtech**

Type: **KT63DL**

Serial No: **B040730**

Date of Calibration: **14-06-2023**

Pre-cal Status: -

Test Conditions

The ambient temperature throughout the period of test was $20 \pm 2^{\circ}\text{C}$ and the relative humidity $50\% \pm 10\% \text{ rh}$.

The instrument was calibrated using Laboratory Standards traceable to BS EN ISO/IEC 17025 Standards.

The calibration and procedures used are proven to meet the requirements of ISO 9001:2015.

All results recorded were found to meet the manufacturer's published specification

No Adjustments were made and unit is functioning correctly at time of calibration

Standards Used: **Transmille 3200 - Fluke 5520A**

Calibrated by: **James Breen**

Certified by: **Ian Park**

The reported uncertainty is based on standard uncertainty multiplied by a covering factor $k=2$, providing a level of confidence of approximately 95%

CERTIFICATE OF CALIBRATION

Certificate No.

0262538

Page 2 of 3

ETHOS 8400 PAGE 2, ETHOS 8400 PAGE 3

Range	Applied Value	Indication	Uncertainty
Ω	1.00 Ω	1.00 Ω	0.5% + 1 digit
	10.00 Ω	10.01 Ω	"
	1.00 kΩ (1000 Ω)	999.50 Ω	"
250V MΩ	0.10 MΩ	0.10 MΩ	"
	90.00 MΩ	89.74 MΩ	"
500V MΩ	0.10 MΩ	0.10 MΩ	"
	90.00 MΩ	89.81 MΩ	"
1000V MΩ	0.10 MΩ	0.10 MΩ	"
	90.00 MΩ	89.76 MΩ	"
HI Loop	1.01 Ω	1.06 Ω	1% + 0.05Ω
	18.75 Ω	18.81 Ω	"
	179.8 Ω	179.70 Ω	"
No Trip Loop	1.01 Ω	1.05 Ω	"
	18.75 Ω	18.79 Ω	"
	179.8 Ω	180.10 Ω	"

Insulation Output Voltages

250V into 0.25MΩ 258 V 500V into 0.50 MΩ 521 V 1000V into 1.00MΩ 1040 V

Uncertainty Of Output Voltages ± 2V

CERTIFICATE OF CALIBRATION

Certificate No.

0262538

Page 3 of 3

ETHOS 8400 PAGE 2, ETHOS 8400 PAGE 3

Range	Applied Value	Indication	Uncertainty
			2.5%
X1 @ 0°	10mA	10.54 mA	"
	20mA	N/A mA	"
	30mA	31.78 mA	"
	100mA	104.1 mA	"
	300mA	310.7 mA	"
	500mA	520.5 mA	"
X5 @ 0°	30mA	159.3 mA	"
X1 @ 180°	30mA	31.79 mA	"
TRIP TIMES			
X1 @ 0°	30mA @ 200 ms	200.4 ms	2ms + 1 lsd
X1 @ 180°	30mA @ 200 ms	200.4 ms	"
X5 @ 0°	30mA @ 30 ms	30.2 ms	"