

Certificate of Calibration

HEME International Ltd

West Lancashire Investment Centre, Maple View,

White Moss Business Park, Skelmersdale,

Lancashire, WN8 9TG, United Kingdom.

TEL: 01695 213154, Email: info@heme-international.com



Certificate No:
CW04330

Date of Calibration
26/06/2026

Approved By :
CALLUM WINTER

Customer Information : RJS SERVICES

Tested Instrument Information :

Instrument Type: TORQUE SCREWDRIVER
Model No.: WG271
Manufacturer: ARMEG
Serial No.: 01-22-A50624
Asset No.:
Work Description :

Traceability Information :

Environmental : TEMPERATURE 20C +/- 3C RELATIVE HUMIDITY 50% +/- 20%

Test Equipments :

Test Equipment Type	Manufacturer	Model No.	Serial No.
TORQUE WRENCH TESTER	NORBAR	TRUCHECK 1-25 NM	82195

Uncertainties :

TORQUE	+/- 2%
--------	--------

Instrument tested to manufacturers specification and EN61557 where applicable

This certificate provides traceability of measurement to recognised National Standards and to the units of measurement realised at the National Physical Laboratory or other National Standards laboratories.

Copyright of this certificate is owned by the issuing laboratory and may not be reproduced except with the prior permission of the issuing laboratory.

This certificate complies with the requirements of BS EN ISO 10012:2003

Certificate of Calibration

Certificate No. CW04330

TEST	APPLIED VALUE	READING	PASS / FAIL
TORQUE (Nm)	1.00	1.04	PASS
	2.00	2.04	PASS
	3.00	2.96	PASS
	4.00	3.95	PASS
	5.00	4.93	PASS
	6.00	5.92	PASS

This certificate provides traceability to recognised National Standards and to the units of measurement realised at the National Physical Laboratory or other recognised National Standards Laboratories.
Copyright of this certificate is owned by the issuing laboratory and may not be reproduced except with the prior written approval of the issuing laboratory.